



MEDISCA[®]
YOUR TRUSTED PARTNER IN COMPOUNDING

Product No.:



PreciseDose Dispenser™ Syringe

DESCRIPTION: The PreciseDose Dispenser™, MEDISCA's line of branded syringes, is an FDA-registered medical device manufactured in an ISO 13485 facility. These oral and topical syringes are designed uniquely for the compounding industry. With quality in mind, these syringes were tested against graduated capacity and leak-resistance using BS/EN ISO 7886-1 – the standard used for sterile hypodermic syringes. MEDISCA has taken a classic design and enhanced it to adhere to modern day oral and topical syringe applications.

FEATURES:

- ✔ For oral and topical use
- ✔ FDA-Registered Medical Device
- ✔ Leak-Resistance and Graduated Capacity tested against BS/EN ISO 7886-1
- ✔ Passed Testing for Ink Smudging
- ✔ Easy-to-read graduated markings on both amber and clear syringes
- ✔ Clear tip caps included with syringe purchase
- ✔ Compatible with MEDISCA Adapter Caps and other bottle adapters
- ✔ Tip caps available in 10 different colors for easy medication identification
- ✔ Available in clear or amber in sizes: 0.5, 1, 3, 5, 10, 20, 35, 60mL
- ✔ PVC, BPA & Latex Free
- ✔ Manufactured in an ISO 13485 facility

AVAILABLE SIZES:

Product No.	Volume	Color
8154	0.5 mL	Amber
8153	1 mL	Amber
8152	3 mL	Amber
8151	5 mL	Amber
8150	10 mL	Amber
8149	20 mL	Amber
7334	35 mL	Amber
7333	60 mL	Amber

Product No.	Volume	Color
8155	0.5 mL	Clear
8148	1 mL	Clear
8147	3 mL	Clear
8146	5 mL	Clear
8145	10 mL	Clear
8144	20 mL	Clear
7336	35 mL	Clear
7335	60 mL	Clear

LIST OF COMPONENTS & MATERIALS:

Part	Material
Barrel	Polypropylene (PP)*
Plunger	Polypropylene (PP)*
Seal	Silicone
Tip Cap	High Density Polyethylene (HDPE)
Lubricant	Polydimethylsiloxane

* FDA approved polypropylene resin, which meets the specifications prescribed in U.S. FDA 21 CFR 177.1520

ADDITIONAL INFORMATION:

Automated Non-Sterile Liquid Filling Process:

- 🔪 7275: Syringe Stand Any MEDISCA PreciseDose Dispenser™
- 🔪 7395: Tubing (IV Spike Inlet to Luer Lock Outlet)
- 🔪 7396: PharmAssist Fluid Transfer Pump
- 🔪 9903: Luer Lock to Luer Slip Connector – Allows PreciseDose Dispenser™ to attach to tubing

Semi-Automated Non-Sterile Cream Filling Process:

- 🔪 6012: Modified 1L Samix Jar – Any sized Samix jar will also work
- 🔪 6664: Pink Nozzle – For front-filling PreciseDose Dispensers™
- 🔪 6668: Samix Applicator, Long – For back-filling PreciseDose Dispensers™
- 🔪 7778: Topi-Click Filling Nozzles – For back-filling PreciseDose Dispensers™
- 🔪 6682: Eazy-Fill™ Topical Transfer Gun
- 🔪 7947: Topi-Click Filling Station
- 🔪 Any MEDISCA PreciseDose Dispenser™

Accuracy Validation:

Tolerance on Graduated Capacity*

The test that we perform for the tolerance on graduated capacity is demonstrative of the syringe’s ability to give an accurate dose of any volume. It establishes the accuracy limits for a specific interval volume that is tested along the length of the barrel.

USA | CANADA | AUSTRALIA PACIFIC | EUROPE
WWW.MEDISCA.COM

Please note that the information contained herein is for informational purposes only. The information is based on recognized compendia, manufacturers’ instructions and/or other authoritative sources and texts. MEDISCA takes no responsibility for the validity or accuracy of this information. Customer MUST refer to USP, FDA, Health Canada, TGA and/or their State Board to ensure they comply with the specific regulations and ensure that appropriate procedures are in place.

Table 1: Accuracy Limits of Target Intervals for Validation of Tolerance on Graduated Capacity for PreciseDose Dispensers™

Nominal Capacity	Target Interval Tested [mL]	Greater than or Equal to Half Nominal Capacity		Below Half Nominal Capacity	
		Lower Limit [mL]	Upper Limit [mL]	Lower Limit [mL]	Upper Limit [mL]
0.5 mL	0.1	0.0950	0.1050	0.0905	0.1095
1 mL	0.1	0.0950	0.1050	0.0830	0.1170
3 mL	0.5	0.525	0.525	0.445	0.555
5 mL	1.0	1.040	1.040	0.915	1.085
10 mL	2.0	2.080	2.080	1.830	2.170
20 mL	5.0	5.200	5.200	4.650	5.350
35 mL	5.0	5.200	5.200	4.425	5.575
60 mL	10.0	10.400	10.400	9.000	11.000

*Derived from the criteria according to BS/EN ISO 7886-1